

## Instructions for Use: Autoclave Tape

Brand Name of Product	Autoclave Tape		
Generic Name of Product	Autoclave Tape		
Product Code Number(s)	ATP-xxx		
Intended Use	Steam autoclave tapes are manufactured for closing packages of items requiring steam		
	sterilization in hospitals a	nd other similar	uses.
Range of Applications for Product	Instrument wraps and oth	er items that need	d to be taped closed during sterilization
Key Specifications of Product		ith an adhesive.	Cator are produced by using strong impregnated This is principally based on natural dry rubber  Typical Value 140-170 9.0 – 10.0 70 - 90 7  Natural, indicator tape printed with white

Shipping & Storage		
<b>Shipping Conditions &amp; Requirements</b>		
Storage Conditions	It is recommended that the products are kept in the original, closed transport cartons and stored in dry and clean conditions protected from direct sunlight and excessive moisture and heat.	
Packaging Contents	<ul> <li>Steam autoclave tapes are packed in recyclable, brown hard carton cases. As an inner cover against moist and dust the steam indicator tapes are wrapped in a recyclable HD polyethylene sheet.</li> <li>Labelling: Each case bears a label with the necessary information/instructions of the content in accordance with ISO 11140-1:2005.</li> </ul>	
Shelf Life		

Instructions for Using Product		
Description of Use (s)	Use to provide easy content identification for instruments or color code items, etc.	
Preparation for Use	Place new roll of tape in the dispenser.	
Diagrams (drawings, pictures)		
Steps for Use of Product	1. Use tape to close packages and items for sterilization.	
	2. Indicator (Class 1) tapes with STEAM indicator changing color from green to brown/black. (The ITSG indicator tapes' color change may vary due to differences in processing parameters.)	
	3. Indicator tapes with EO gas indicator changing color from brown to green.	
Interpretation of Results		
<b>Contraindications of Test Results</b>		
Documentation		
Special Warnings and Cautions	Tape is single-use only	
	The indicators should not be exposed to direct light prior to use.	
Disposal		

Reprocessing Instructions		
Point of Use		
Preparation for Decontamination		
Disassembly Instructions		
Cleaning – Manual		
Cleaning – Automated		
Disinfection		
Drying		
Maintenance, Inspection, and Testing		

Reassembly Instruction	
Packaging	
Sterilization	Steam The steam indicator ink will react positively when subjected to heat and moisture in the steam autoclave. After sterilization strong dark lines appear on the backing.  • Pre-vac at 4 minutes/270°F, 3 minutes at 275°F  • Gravity at 15 minutes at 250°F.  EO The ethylene oxide indicator ink will react positively when subjected to ethylene oxide
	autoclave process.
Storage	
Additional Information	

Related Healthmark Products		
Other Product Support Documents	Labeling Brochure, Labeling Price List	
Reference Documents		
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc.	
	18600 Malyn Blvd.	
	Fraser, MI 48026	
	1-586-774-7600	
	healthmark@hmark.com	
	hmark.com	