## CartWashCheck™ Test Log

Make/	Model	of the	Medical	Cart Was	her
#					

Date Tested	Testers Initials	Item Tested	Cycle Selection	Temperature Result	Water Exposure Result	Action Taken	Other
	avotiono:						

Observations:	 	 

<sup>\*</sup>Record all results: correct result should be white background and a minimum of 120°F; on CartWashCheck™. Please note that 120°F might not be the temperatures required at each facility consult your management and manufacture of the equipment for correct temperature for the cycle selected. The special hydrophilic ink will change color (from black/dark blue to white) only if moistened by water. This demonstrates that water is reaching the area tested and indicates proper mechanical action by the washer. White Means Water Reached the Test. The square needs to be completely white no black/dark blue color left on the square.