

## Instructions for Use: Sample Extraction Tools

<b>Brand Name of Product</b>	Sample Extraction Tool
<b>Generic Name of Product</b>	Single use sample extraction brushes
<b>Product Code Number(s)</b>	CKB-200-200, CKB-230-400, CKB-230-600
<b>Intended Use</b>	The sample extraction tool is intended for extracting samples to test for bacterial growth.
<b>Range of Applications for Product</b>	Extracting samples for testing.
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• Proper length and diameters for contact with the internal surfaces.             <ul style="list-style-type: none"> <li>- CKB 200-200: 200cm Length x 2mm Diameter</li> <li>- CKB 230-400: 230cm Length x 4mm Diameter</li> <li>- CKB 230-600: 230cm Length x 6mm Diameter</li> </ul> </li> </ul>

Shipping & Storage	
<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Contents</b>	
<b>Shelf Life</b>	

Instructions for Using Product	
<b>Description of Use(s)</b>	To extract samples for testing bacterial growth.
<b>Preparation for Use</b>	<ol style="list-style-type: none"> <li>1. Select the diameter and length for the internal surface/channel.</li> <li>2. Remove the extraction tool from the packaging.</li> </ol>
<b>Diagrams (drawings, pictures)</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Insert the sample extraction tool into the channel.</li> <li>2. Continue to feed the extraction tool through the length of the channel until it reaches the distal end. Never force the extraction tool, to avoid possible damage.</li> <li>3. Once the extraction tool exits the distal end, cut the entire head of the bristle portion of the extraction tool into the sample collection container.</li> <li>4. Carefully pull the entire sample extraction tool completely out of the distal end.</li> <li>5. Dispose of the extraction tool after use.</li> </ol>
<b>Interpretation of Test Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	<ul style="list-style-type: none"> <li>• Select the extraction tool based upon the diameter and length of the channel chosen.</li> <li>• Do not force the extraction tool into a channel to avoid possible damage.</li> </ul>
<b>Disposal</b>	Treat used sample extraction tools as a biohazard and dispose of in accordance with facility policy of contaminated items.

Reprocessing Instructions	
<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	
<b>Reassembly Instructions</b>	
<b>Packaging</b>	
<b>Sterilization</b>	
<b>Storage</b>	

<b>Additional Information</b>	
<b>Related Healthmark Products</b>	Flexible Endoscope Sampling Kits
<b>Other Product Support Documents</b>	ProFormance™ Brochure, ProFormance™ Price List
<b>Reference Documents</b>	
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>