

Instructions for Use: Staph-Check #20 Cart Cover

Brand Name of Product	Staph-Check #20 Cart Cover
Generic Name of Product	Cart Cover
Product Code Number(s)	Staph-Check #20
Intended Use	Ideal for transport carts with an extra level of protection
Range of Applications for Product	Medical supplies and accessories
Key Specifications of Product	Staph- Check #20 Dimensions <ul style="list-style-type: none"> • Flame resistant • Bacteria static • Self-deodorizing

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	
Packaging Contents	
Shelf Life	

Instructions for Using Product	
Description of Use(s)	Ideal for transport carts with an extra level of protection.
Preparation for Use	
Diagrams (drawings, pictures)	
Steps for Use of Product	Place cover over the cart so it is completely covered.
Interpretation of Test Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	

Reprocessing Instructions																										
Point of Use																										
Preparation for Decontamination																										
Disassembly Instructions																										
Cleaning – Manual	Hard to clean spots- use a soft sponge with a liquid household/vinyl cleaner. Do not use harsh cleaners or solvents.																									
Cleaning – Automated	<p>A slow speed machine must be used. The wash wheel is to be stopped for draining and refilling between each.</p> <table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="text-align: left;">Operation Cycle</th> <th style="text-align: left;">Water Level</th> <th style="text-align: left;">Time</th> <th style="text-align: left;">Maximum Water Temperature</th> <th style="text-align: left;">Supplies</th> </tr> </thead> <tbody> <tr> <td>1st Suds</td> <td>High</td> <td>5 min.</td> <td>100° F</td> <td>Neutral Soap & Alkali</td> </tr> <tr> <td>2nd Suds</td> <td>High</td> <td>10 min.</td> <td>100° F</td> <td>Neutral Soap</td> </tr> <tr> <td>1st Rinse</td> <td>High</td> <td>3 min.</td> <td>100° F</td> <td>Water</td> </tr> <tr> <td>2nd Rinse</td> <td>Low</td> <td>3 min.</td> <td>90° F</td> <td>Sour to ph 5.0</td> </tr> </tbody> </table>	Operation Cycle	Water Level	Time	Maximum Water Temperature	Supplies	1 st Suds	High	5 min.	100° F	Neutral Soap & Alkali	2 nd Suds	High	10 min.	100° F	Neutral Soap	1 st Rinse	High	3 min.	100° F	Water	2 nd Rinse	Low	3 min.	90° F	Sour to ph 5.0
Operation Cycle	Water Level	Time	Maximum Water Temperature	Supplies																						
1 st Suds	High	5 min.	100° F	Neutral Soap & Alkali																						
2 nd Suds	High	10 min.	100° F	Neutral Soap																						
1 st Rinse	High	3 min.	100° F	Water																						
2 nd Rinse	Low	3 min.	90° F	Sour to ph 5.0																						
Disinfection																										
Drying	<p>Air dry. If machine drying is necessary, tumble dry without heat. Ironing- do not put covers through flatwork ironers.</p> <p>Caution: Washing or drying in excess of the above recommendations will reduce service life and performance of the product.</p>																									
Maintenance, Inspection, and Testing																										

Reassembly Instructions	
Packaging	
Sterilization	
Storage	
Additional Information	
Related Healthmark Products	
Other Product Support Documents	Bins Galore Brochure, Bins Galore Price List
Reference Documents	
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com