

Instructions for Use: Sterimarker

Brand Name of Product	Sterimarker
Generic Name of Product	Sterimarker
Product Code Number(s)	MAR1, MAR2, MAR3, 1753DZ
Intent of Use	To write on packages, in accordance with the IFU of the packaging material.
Range of Applications for Product	The special ink is immediately dry, waterproof, lightproof, resistant to atmospheric action, heat resistant, and acid resistant. No chemical or toxic substances are released during use. The ink marking withstands sterilization by steam, EO, VH2O2 and radiation.
Key Specifications of Product	<ul style="list-style-type: none"> • Nontoxic • Waterproof • No chemicals or toxic substances are released during use.

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	<ul style="list-style-type: none"> • No special precautions are necessary regarding storage/transportation. • Do not expose them to very high humidity or extreme heat.
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use(s)	To write on packages in accordance with the IFU of the packaging material.
Preparation for Use	N/A
Diagrams (drawings, pictures)	N/A
Steps for Use of Product	<ol style="list-style-type: none"> 1. Uncap marker. 2. Write on surfaces identified as appropriate to write on, according to the product's instructions. 3. Recap marker after use.
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	<ul style="list-style-type: none"> • Recap marker immediately after use. • The marker pen is not autoclavable. • Do not write on surfaces with a glossy finish, the ink will smudge after sterilization.
Disposal	N/A

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	N/A
Cleaning – Automated	N/A
Disinfection	N/A
Drying	N/A
Maintenance, Inspection, and Testing	N/A
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	Ink is compatible with Steam sterilization: <ul style="list-style-type: none"> • 132 °C (270 °F) for 4 minutes • 135 °C (275 °F) for 3 minutes • VH2O2/O3/EO sterilization and radiation.
Storage	N/A

Additional Information	N/A
Related Healthmark Products	N/A
Other Product Support Documents	N/A
Reference Documents	Sterilization Brochure, Sterilization Price List
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com