

## Instructions for Use: Instrument Protection Foam

<b>Brand Name of Product</b>	Instrument Protection Foam
<b>Generic Name of Product</b>	Instrument Protection Foam
<b>Product Code Number(s)</b>	FP-4600, FP-4800, FP-4850, FP-4880, FP-5200, FP-5500, FP-6200, FP-6210, FP-6250, FP-6270
<b>Purpose of Product</b>	Single-use foam pouches designed to prevent costly damage to instrument tips caused by shifting during handling and transport. The pouches neither retain nor absorb moisture.
<b>Range of Applications for Product</b>	<ul style="list-style-type: none"> <li>• Small instruments that require tip protection during transport, storage, and sterilization.</li> <li>• Can be used in sterile barrier systems validated for use in steam and EO sterilization.</li> </ul>
<b>Key Specifications of Product</b>	Single-use foam pouches.

Shipping & Storage	
<b>Shipping Conditions &amp; Requirements</b>	N/A
<b>Storage Conditions</b>	N/A
<b>Packaging Contents</b>	N/A
<b>Shelf Life</b>	N/A

Instructions for Using Product	
<b>Description of Use(s)</b>	Single-use foam pouches designed to prevent costly damage to instrument tips caused by shifting during handling and transport.
<b>Preparation for Use</b>	If pouches are rolled, unroll so that the pouches are facing up.
<b>Diagrams (drawings, pictures)</b>	N/A
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Place instrument inside pouch.</li> <li>2. Be sure the tips are secure within the pouch.</li> </ol>
<b>Interpretation of Results</b>	N/A
<b>Contraindications of Test Results</b>	N/A
<b>Documentation</b>	N/A
<b>Special Warnings and Cautions</b>	N/A
<b>Disposal</b>	Single use only.

Reprocessing Instructions	
<b>Point of Use</b>	N/A
<b>Preparation for Decontamination</b>	N/A
<b>Disassembly Instructions</b>	N/A
<b>Cleaning – Manual</b>	N/A
<b>Cleaning – Automated</b>	N/A
<b>Disinfection</b>	N/A
<b>Drying</b>	N/A
<b>Maintenance, Inspection and Testing</b>	N/A
<b>Reassembly Instructions</b>	N/A
<b>Packaging</b>	N/A
<b>Sterilization</b>	<ul style="list-style-type: none"> <li>• For use in sterile barrier systems validated for use in steam and EO sterilization.</li> <li>• Steam sterilization: <ul style="list-style-type: none"> <li>○ 270 °F (132 °C) for 4 minutes</li> <li>○ 275 °F (135 °C) for 3 minutes</li> <li>○ Maximum extended time: 275 °F (135 °C) for 30 minutes.</li> </ul> </li> <li>• Compatible with EO sterilization.</li> </ul>
<b>Storage</b>	N/A
<b>Additional Information</b>	N/A

<b>Related Healthmark Products</b>	Foam Tip Protectors, Foam Corner Protectors, Foam Pinhole Protector
<b>Other Product Support Documents</b>	ProSys™ Brochure, ProSys™ Price List
<b>Reference Documents</b>	N/A
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>