








## Instructions for Use: Spray Gun

<b>Brand Name of Product</b>	Spray Gun
<b>Generic Name of Product</b>	Spray Gun
<b>Product Code Number(s)</b>	6020, 80-00-44, 80-10-15, 4240, 4250, 5220, 5110, 5020, 5030, 1HGL4, 4270, 4260, 5070, 4230, 5040, 6000, 6000-6025, 5120, 5010, 6025, 6045, 14RVT, 5210
<b>Purpose of Product</b>	A tool to use with water or air in the cleaning of items.
<b>Range of Applications for Product</b>	N/A
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>Use with the “universal” slip-tip, or the optional quick connect set of tips for specialized functions.</li> <li>Maximum water or air pressure six (6) bar (86 psi).</li> </ul>
	N/A

Shipping & Storage	
<b>Shipping Conditions &amp; Requirements</b>	N/A
<b>Storage Conditions</b>	N/A
<b>Packaging Conditions</b>	N/A
<b>Shelf Life</b>	N/A

Instructions for Using Product	
<b>Description of Use(s)</b>	A tool to use in the cleaning of items.
<b>Preparation for Use</b>	<ul style="list-style-type: none"> <li>Best installed by a licensed plumber. Please contact Facilities and/or Engineering for installation of the Spray Gun. Installation of the Spray Gun is performed by the healthcare facility.</li> <li>Connection requirements               <ul style="list-style-type: none"> <li><i>Water connection:</i> <ul style="list-style-type: none"> <li>Plumbing threaded tap or valve: 3/8-, 1/2-, or 3/4-inch</li> <li>Standard gun kit includes 3/4-inch female connection.</li> </ul> </li> <li><i>Air connection:</i> If operated with compressed air, the tubing must be connected to the air supply by a suitable connector. (Note: cannot be supplied by Healthmark.)</li> <li><i>Water or air connection below the counter:</i> <ul style="list-style-type: none"> <li>It is recommended that a rosette be installed. It will minimize potential friction damage to the hose.</li> <li>A 1-3/8 inches diameter hole will need to be drilled in the counter or sink top to accommodate the rosette.</li> </ul> </li> </ul> </li> <li>(Prior to operation), the user must make sure attachments are correctly seated on the safety cone.</li> <li>(If the Spray Gun is not in use), ensure the water or air pressure have been turned off.</li> <li>When the hose of the Spray Gun is installed to the water line, a separate shutoff valve needs to be installed to the spigot. (Note: The Spray Gun is not a controller.)</li> </ul>
<b>Diagrams (drawings, pictures)</b>	<div>  <p>Figure 1: 6000-6025 Quick Connect Tips</p> </div> <div>  <p>Figure 2: 5220 Specialized Tips</p> </div>

	 <p>Figure 3: 3000, Nylon Spray Gun Wall Mount Figure 4: 3005, Chrome Brass Spray Gun Wall Mount</p>	 <p>Figure 5: 5210 Spray Gun Brush Set- 7 Brush Attachments and Holder</p>
	 <p>Figure 6: Threaded nut for controlling maximum flow</p>	 <p>Figure 7: Flow rate controlled by trigger handle</p>
	 <p>Figure 8: Diagram Spray Gun Construction</p>	<p><b><u>Key to Spray Gun Components</u></b></p> <ol style="list-style-type: none"> <li>1. Gun Rinser Tip Holder complete kit</li> <li>2. "O" Ring black 3 mm x 1 mm</li> <li>3. Spiral Tension Pin</li> <li>4. Trigger Handle</li> <li>5. Grip for Spray Gun</li> <li>6. Hose Connector</li> <li>7. Gun Cap</li> <li>8. Flow Control Kit</li> </ol>
<p><b>Steps for Use of Product</b></p>	<p>Prior to operation, the user must:</p> <ol style="list-style-type: none"> <li>1. Take care not to point Spray Gun at self/others' orifices prior to operation.</li> <li>2. Make sure attachments are correctly seated on the safety cone. <ol style="list-style-type: none"> <li>a. Firmly push the attachments onto the Spray Gun's rinse tip.</li> <li>b. Ensure the hose's valve is open.</li> </ol> </li> </ol> <p>When in operation:</p> <ol style="list-style-type: none"> <li>3. Wear appropriate PPE per the facility's policy.</li> <li>4. Set Spray Gun's maximum pressure with the small threaded nut located before the trigger handle.</li> </ol>	
<p><b>Interpretation of Results</b></p>	<p>N/A</p>	
<p><b>Contraindications of Test Results</b></p>	<p>N/A</p>	
<p><b>Documentation</b></p>	<p>N/A</p>	
<p><b>Special Warnings and Cautions</b></p>	<p>Cautions:</p> <ul style="list-style-type: none"> <li>• Verify the safe and effective performance of any tubing chosen for use other than that supplied by Healthmark. (Note: The liability for damage or injury is the responsibility of the facility.)</li> <li>• Warranty claims will <i>only</i> be honored for components purchased from Healthmark.</li> <li>• Maximum water or air pressure to use is six (6) bar (86 psi).</li> <li>• Be sure to wear proper PPE according to facility policy.</li> </ul> <p>Warnings:</p> <ul style="list-style-type: none"> <li>• ⚠ Do not use Spray Gun for any medical treatments.</li> <li>• ⚠ Do not point the Spray Gun at parts or orifices of the body when in use. Compressed water or air jet may be harmful.</li> </ul>	

<b>Disposal</b>	N/A
-----------------	-----

Reprocessing Instructions	
<b>Point of Use</b>	N/A
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	N/A
<b>Cleaning – Manual</b>	Use warm water <i>only</i> and a mild detergent to clean the Spray Gun with all add-ons.
<b>Cleaning – Automated</b>	N/A
<b>Disinfection</b>	N/A
<b>Drying</b>	N/A
<b>Maintenance, Inspection, and Testing</b>	<ul style="list-style-type: none"> <li>● <i>Recommendation:</i> decalcify the Spray Gun approximately every four (4) weeks. <ul style="list-style-type: none"> <li>○ Beneficial for the service life and reliability of the Spray Gun.</li> <li>○ Disassemble the Spray Gun.</li> <li>○ Place the inner-parts in a decalcifying agent.</li> </ul> </li> <li>● It is not necessary to carry out water quality tests when using potable water.</li> <li>● Replacement parts and instructions for repair are available from Healthmark.</li> </ul>
<b>Reassembly Instructions</b>	<ul style="list-style-type: none"> <li>● The complete reassembly of the clean gun must be carried out under sterile conditions.</li> <li>● Ensure that all O-Rings are in perfect condition before screwing the individual parts together.</li> <li>● (<i>Must</i>) moisten sealing elements with a suitable physiological saline solution before use.</li> <li>● Mount the sealing rings at the intended locations and lock them in place by pressing them firmly by hand.</li> </ul>
<b>Packaging</b>	N/A
<b>Sterilization</b>	N/A
<b>Storage</b>	N/A
<b>Additional Information</b>	N/A

<b>Related Healthmark Products</b>	N/A
<b>Other Product Support Documents</b>	ProSys™ Brochure, ProSys™ Price List
<b>Reference Documents</b>	N/A
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> hmark.com