



Instructions for Use: Translucent Tip Protectors

Brand Name of Product	Translucent Tip Protectors
Generic Name of Product	Tip Protectors
Product Code Number(s)	10001, 10002, 10003, 10004, 10005, 10006, 10007, 10008, 10009, 10010
Intended Use	Prevents moisture from being trapped between the protector and the item. Tip Protectors are also vented facilitating evaporation during the sterilization cycle. The see-through design means tips are easily visible.
Range of Applications for Product	Items with delicate tips.
Key Specifications of Product	<ul style="list-style-type: none"> • Compatible with steam and EO sterilization • Ribbed design which facilitates sterilant circulation and reduces the chance for sticking to instruments • Vented holes to facilitate sterilant circulation • Translucent material for visibility • Available in 10 sizes • Manufactured of heat stable TPE • Single-use

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	
Package Contents	100 per box
Shelf Life	

Instructions for Using Product	
Description of Use(s)	To be placed on items during sterilization to protect the delicate and sharp tips.
Preparation for Use	Fit each item with the appropriate size tip.
Diagrams (drawings, pictures)	
Steps for Use of Product	<ul style="list-style-type: none"> • Select the appropriate tip protector for the tip that is to be protected. • Package in sterile barrier packaging as directed by the item's IFU. • Place package in the appropriate sterilizer as directed by the item's and sterile barrier system manufacturer's IFU.
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	Dispose of in compliance with facility guidelines.

Reprocessing Instructions	
Point of Use	
Preparation for Decontamination	
Disassembly Instructions	
Cleaning – Manual	
Cleaning – Automated	
Disinfection	
Drying	
Maintenance, Inspection, and Testing	
Reassembly Instructions	
Packaging	
Sterilization	<ul style="list-style-type: none"> • Recommended steam sterilization cycles are 4 minutes at 132°C or 3 minutes at 135°C in a pre-vac sterilizer. Gravity at 121°C for 30 minutes • EO at 55°C, 60 minutes, 12 hours aeration time
Storage:	
Additional Information	

Related Healthmark Products	Tip Protectors, Vented Instrument Guards, Duo-Vented Instrument Guards, Cylindrical Style, Instrument Guards.
Other Product Support Documents	http://www.healthmark.info/InstrumentCare/TipProtectors/Translucent_Tip_Protector_Study_Summary.pdf
Reference Documents	
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com