



Instructions for Use: Scope Holder

Brand Name of Product	Scope Holder
Generic Name of Product	Scope Holder
Product Code Number(s)	SCPH-102
Intended Use	Stores flexible endoscopes upright.
Range of Applications for Product	Storage for endoscopes
Key specifications of product	<ul style="list-style-type: none"> • Plexiglas® wall mount

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Conditions	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use (s)	Stores flexible endoscopes upright
Preparation for Use	N/A
Diagrams (drawings, pictures):	N/A
Steps for Use of Product	<ol style="list-style-type: none"> 1. Once scope has been reprocessed, hang in an upright position in accordance with the scope manufacture IFU. 2. Keep stored until being evaluated before next use.
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	<ul style="list-style-type: none"> • Mounting to typical wall with 2”x 4” studs, there will be a thin side of the stud, approximately 1.75” to anchor to. The holes will not allow for all four to be anchored on a stud this size. It is NOT advisable to anchor this directly to drywall as it would probably fall off.
Disposal	N/A

Reprocessing Instructions	
Point of use:	N/A
Preparation for decontamination:	N/A
Disassembly Instructions:	N/A
Cleaning – Manual:	Detergent and warm water.
Cleaning – Automated:	N/A
Disinfection:	N/A
Drying:	N/A
Maintenance, inspection, and testing:	N/A
Reassembly Instructions:	N/A
Packaging:	N/A
Sterilization:	N/A
Storage:	N/A
Additional Information:	N/A

Related Healthmark Products	N/A
------------------------------------	-----

Other Product Support Documents	ProSys™ Price Brochure, ProSys™ Price List
Reference Documents	N/A
Customer Service contact:	Healthmark Industries Company, Inc 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com