

Department	
------------	--

Test Date	Tester Initials	Endoscope Model Tested	Endoscope #	Pass/Fail*	Comments

<b>PASS</b>	<b>NO</b> color change on the swab
<b>FAIL</b>	<b>A COLOR CHANGE TO PURPLE</b> on the swab indicates residual moisture at or above the level of detection