




Instructions for Use: Tray Identification Tags

Brand Name of Product	Tray Identification Tags
Generic Name of Product	Tray Identification Tags
Product Code Number(s)	RING001, TG001, TG001-E, TG011, TG012, TG013, TG014, TG015, TG016, TG017, TG018, TG019, TG020, TG021, TG022, TG023, TG024, TG025, TG026, TG011-E, TG012-E, TG013-E, TG014-E, TG015-E, TG016-E, TG017-E, TG018-E, TG019-E, TG020-E, TG021-E, TG022-E, TG023-E, TG024-E, TG025-E, TG026-E, TIE003, STT-001,
Intended Use	For identification, tracking, and tracing purposes.
Range of Applications for Product	A range of colored plastic labeling tags for identification/tracking of trays, baskets, and other items.
Key Specifications of Product	<ul style="list-style-type: none"> • Resin bonded. • Various colors. • X-ray detectable. • Autoclavable.

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use(s)	For use with trays, baskets, and other items for identification purposes.
Preparation for Use	N/A
Diagrams (drawings, pictures):	 
Steps for Use of Product	HOW TO USE TRAY TAGS

	<p>(Note: Care should be taken to ensure the writing (with a suitable marking pen) will not be removed or bleed during processing. We recommend to test this prior to placing in use.)</p> <p>Writing directly onto the Tray Identification Tag:</p> <ol style="list-style-type: none"> Information may be written directly on the Tag. Either or both sides of the Tag may be written on: <ol style="list-style-type: none"> Using a pre-printed self-adhesive label Pre-printed self-adhesive labels may be stuck onto either side of a Tag. <p>When Attaching the Tray Identification Tag to Tray or Basket or other items:</p> <ol style="list-style-type: none"> Use the holes in the Tag to attach it to the tray or basket. Use a Split Ring (RIN001) or a Security Tie (TIE003) for this purpose. If X-ray detectability is required, use a Split Ring (RIN001) to attach the Tag to the tray since the split ring is X-ray detectable.
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	<ul style="list-style-type: none"> Facilities should develop a policy for the placement of the Tray Identification Tag on the tray or basket. This should include a risk assessment to avoid potential interference with effective processing or aseptic presentation. Store away from any sources of ignition. 
Disposal	<ul style="list-style-type: none"> Users should reference facility procedures. Unused Tray Identification Tags may be disposed of in the plastic waste receptacle. Used Tray Identification Tags should be disposed of by considering possible microbial contamination.

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	N/A
Cleaning – Automated	Tray Identification Tags are compatible with washer disinfectors conforming to ISO 15883 series of standards.
Disinfection	N/A
Drying	N/A
Maintenance, Inspection, and Testing	N/A
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	<ul style="list-style-type: none"> Materially compatible with Steam, EO, and VH2O2 Sterilization. Recommended cycle for steam sterilization: <ul style="list-style-type: none"> 132 °C (270 °F) for four (4) minutes 135 °C (275 °F) for three (3) minutes. Extended steam sterilization cycle: 135 °C (275 °F) for 30 minutes.
Storage	N/A
Additional Information	N/A

Related Healthmark Products	N/A
Other Product Support Documents	Labeling Brochure, Labeling Price List
Reference Documents	N/A
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com