

Instructions for Use: Cable-Tie Style Loks

Brand Name of Product	Cable-Tie Style Loks
Generic Name of Product	Tamper Evident Seal
Product Code Number(s)	8525 WHITE
Purpose of Product	To show tampering and/or access to critical equipment and supplies.
Range of Applications for Product	Crash carts, emergency boxes, cabinets, and other areas where demonstrating tampering is
	critical.
Key Specifications of Product	Numbered
	 Cynch-tight nylon security straps with a steel locking tooth
	Seal and broken by hand
	Cannot be resealed once broken

Shipping & Storage		
Shipping Conditions & Requirements		
Storage Conditions		
Packaging Contents		
Shelf Life		

Instructions for Using Product	
Description of Use (s)	Seal crash carts, emergency boxes and other areas where evidence of tampering is critical
	when a longer seal is needed
Preparation for Use	
Diagrams (drawings, pictures)	
Steps for Use of Product	1. Insert seal into closing of two sides of the object being secured.
	2. Seal through the locking dome of the seal.
	3. Pull tail of seal as tight as possible to ensure item is tightly sealed.
	4. The seal can be easily opened with a tug.
Interpretation of Results	
Contraindications of Test Results	
Documentation	Follow the policy of your facility for recording date, staff and other information critical to
	management of critical inventory.
Special Warnings and Cautions	If a seal is broken, assume items have been tampered with or removed. Follow your facility
	instructions for inspection and replacement of critical inventory.
Disposal	

Reprocessing Instructions	
Point of Use	
Preparation for Decontamination	
Disassembly Instructions	
Cleaning – Manual	
Cleaning – Automated	
Disinfection	
Drying	
Maintenance, Inspection, and Testing	
Reassembly Instructions	
Packaging	
Sterilization	Compatible with steam sterilization. Standard cycle: 270 °F/4 minutes
Storage	
Additional Information	

Related Healthmark Products	
Other Product Support Documents	Security Products Brochure, Security Products Price List
Reference Documents	
Customer Service Contact	Healthmark Industries Company, Inc
	18600 Malyn Blvd.

Fraser, MI 48026
1-586-774-7600
healthmark@hmark.com
hmark.com